

This is the peer reviewed version of the following article:

Oral surgery biomaterials: analyses of Al₂O₃-treated titanium surfaces tested with fibroblast and osteocyte cell lines / Smargiassi, Alberto; Ferretti, Marzia; Cavani, Francesco; Sena, Paola; Benincasa, Marta; Zaffe, Davide; Facciani, Valentino; Gabrel, Ivano; Palumbo, Carla. - In: ITALIAN JOURNAL OF ANATOMY AND EMBRYOLOGY. - ISSN 1122-6714. - STAMPA. - 119 - n. 1 (suppl.):(2014), pp. 184-184. (Intervento presentato al convegno 68th Meeting of the Italian Society of anatomy and Histology tenutosi a Ancona (Italy) nel 18-20 September 2014).

Firenze University Press

Terms of use:

The terms and conditions for the reuse of this version of the manuscript are specified in the publishing policy. For all terms of use and more information see the publisher's website.

12/10/2024 02:08

(Article begins on next page)

Oral surgery biomaterials: analyses of Al₂O₃-treated titanium surfaces tested with fibroblast and osteocyte cell lines

Alberto Smargiassi¹, Marzia Ferretti¹, Francesco Cavani¹, Paola Sena¹, Marta Benincasa¹, Davide Zaffe¹, Valentino Facciani², Ivano Gabrel², Carla Palumbo¹

¹Dipartimento di Scienze biomediche, Metaboliche e Neuroscienze - Sez. Morfologia umana, Univ. Modena-RE, IT

²SAFE & SIMPLE, Treviso, IT

Two different cell lines - MLO-Y4 (murine osteocytes) and 293 (human fibroblasts) - cultured for 48 hours in standard media were used to analyse engineered biomaterials (i.e. Al₂O₃ shot-peened titanium surfaces). Distribution, density and expression of adhesion molecules (fibronectin and vitronectin) were evaluated under scanning electron microscope (SEM) and confocal microscope (CM) as previously described [1]. The engineered biomaterial surfaces showed under SEM irregular morphology displaying variously-shaped spicules, obtained by shooting different-in-size particles of Al₂O₃ against the scaffolds of biomaterial. DAPI and fluorochrome-conjugated antibodies were used to highlight nuclei, fibronectin and vitronectin, under CM; cell distribution was analysed after Gold-Palladium sputtering of samples by SEM. Both SEM and CM observations showed better outcome in terms of cell adhesion and distribution in treated titanium surfaces with respect to the untreated ones. The results obtained clearly showed that this kind of surface-treated titanium, used to manufacture devices for dental implantology: *i*) is very suitable for cell colonization, essential prerequisite for the best osseointegration, and *ii*) represents an excellent solution for the development of further engineered implants with the target to obtain recovery of dental function stable over time.

Further studies on these Al₂O₃ shot-peened-titanium surfaces, both *in vitro* and *in vivo*, will be needed to obtain accurate definition of better biomaterial outcome, also after additional treatments.

References

[1] Palumbo et al. (2013) Immunocytochemical and structural comparative study of committed *versus* multipotent stem cells cultured with different biomaterials. *Micron* 47: 1-9.

Keywords

Titanium scaffold, MLO-Y4 and 293 cultures, immunofluorescence, confocal microscopy, scanning electron microscopy.