

This is the peer reviewed version of the following article:

HyFlex EDM rotary Ni-Ti prototypes: the effect of an innovative machining technology on Ni-Ti wear /
Pirani, C.; Generali, Luigi; Iacono, F.; Sassetelli, Paolo; Prati, C.. - In: MINERVA STOMATOLOGICA. - ISSN
0026-4970. - ELETTRONICO. - 64:(2015), pp. 51-51. (Intervento presentato al convegno XXII Congresso
Nazionale Collegio dei Docenti Universitari di Discipline Odontostomatologiche tenutosi a Milano nel
9,10,11 aprile 2015).

Collegio dei Docenti di Odontoiatria

Terms of use:

The terms and conditions for the reuse of this version of the manuscript are specified in the publishing policy. For all terms of use and more information see the publisher's website.

19/04/2024 04:15


(Article begins on next page)

MINERVA STOMATOLOGICA

VOLUME 64 · SUPPL. I · N. 2 · APRIL 2015

**XXII Congresso Nazionale
Collegio dei Docenti Universitari
di Discipline Odontostomatologiche**

Milano, 9-11 aprile 2015



EDIZIONI · MINERVA · MEDICA

MINERVA STOMATOLOGICA

RIVISTA BIMESTRALE DI ODONTOSTOMATOLOGIA FONDATA NEL 1952
DA B. ROCCIA E T. OLARIO

continuazione di: MINERVA ORTOGNATODONTICA
ORGANO UFFICIALE DELLA SOCIETÀ ITALIANA
DI ODONTOSTOMATOLOGIA E CHIRURGIA MAXILLO-FACCIALE

REDAUTTORE CAPO
C. Paganelli

DIRETTORE SCIENTIFICO
P. L. Sapelli

COMITATO DI REDAZIONE

Biglioli F., Borracchini A., Brambilla E., Breschi L., Campisi G., Carossa S., Carozzo M., Carmagnola D., Cerutti A., Chiapasco M., Cicciù D., Cicciù M., Cordasco G., Cortellini P., D'Aiuto F., Dalessandri D., Dalessandri M., Del Frabbro M., Di Rosario F., Dorigo E., Esposito M., Favia G., Ferrari M., Ferrario V., Fontana P., Francetti L., Gaffuri S., Gagliani M., Gastaldi G., Gherlone E., Laffranchi L., Levriani L., Lo Muzio L., Lodi G., Maggiore C., Maiorana C., Martina R., Montebugnoli L., Mortellaro C., Paganelli C., Pelo S., Perfetti G., Pini Prato G., Pizzo G., Prati C., Putignano A., Rimondini L., Rocuzzo M., Romeo E., Salgarello S., Sammartino G., Sanz M., Sardella A., Scarano A., Scully C., Sesenna E., Tian C., Tonni I., Trombelli L., Urbani G., Van Der Waal I., Warnakulasuriya S.

CONSIGLIO DIRETTIVO S.I.O.C.M.F.

Presidente
D. Cicciù

Vice Presidenti
C. Maggiore, C. Paganelli

Segretario generale
E. Rapisarda

Segretario Tesoriere
M. Procaccini

Consiglieri

M. Bosco, S. Caputi, A. Carrassi, U. Covani, C. De Bernardi,
G. P. Farronato, E. Gherlone, C. Mortellaro, A. Polimeni, F. Zarone

Revisori dei Conti

A. Forabosco, G. Gallina, S. Pizzi

Past President - Probi Viri

E. Bertelli, R. Grandini, G. Nidoli, S. Pelo, P. L. Sapelli

DIRETTORE RESPONSABILE

A. Oliaro



This journal is PEER REVIEWED and is quoted in: CAB, EMBASE, Index to Dental Literature, PubMed/MEDLINE, Scopus

Direzione, redazione, ufficio grafico, ufficio pubblicità, fotocomposizione, amministrazione - Edizioni Minerva Medica - Corso Bramante 83-85 - 10126 Torino
Tel. (011) 67.82.82 - Fax (011) 67.45.02 - E-mail: minervamedica@minervamedica.it
Web Site: www.minervamedica.it

Stampa - Edizioni Minerva Medica - Tipografia di Saluzzo - Corso IV Novembre 29-31 - 12037 Saluzzo (CN) - Tel. (0175) 249405 - Fax (0175) 249407
Italia - Individuale: Online € 110,00, Cartaceo € 115,00, Cartaceo+Online € 120,00; **Istituzionale:** Cartaceo € 150,00, Online (Small € 282,00, Medium € 325,00, Large € 373,00, Extra Large € 390,00), Cartaceo+Online (Small € 294,00, Medium € 341,00, Large € 388,00, Extra Large € 404,00).

Unione Europea - Individuale: Online € 185,00, Cartaceo € 190,00, Cartaceo+Online € 200,00; **Istituzionale:** Cartaceo € 265,00, Online (Small € 285,00, Medium € 325,00, Large € 373,00, Extra Large € 390,00), Cartaceo+Online (Small € 304,00, Medium € 351,00, Large € 399,00, Extra Large € 414,00).

Paesi extraeuropei - Individuale: Online € 205,00, Cartaceo € 215,00, Cartaceo+Online € 225,00; **Istituzionale:** Cartaceo € 290,00, Online (Small € 304,00, Medium € 341,00, Large € 393,00, Extra Large € 409,00), Cartaceo+Online (Small € 320,00, Medium € 362,00, Large € 414,00, Extra Large € 430,00).

Per il pagamento: 1) Gli abbonati possono utilizzare le seguenti forme di pagamento: a) assegno bancario; b) bonifico bancario a: Edizioni Minerva Medica, INTESA SANPAOLO Agenzia n. 18 Torino, IBAN: IT45 K030 6909 2191 0000 0002 917 c) conto corrente postale 00279109 intestato a Edizioni Minerva Medica, Corso Bramante 83-85, 10126 Torino; d) carte di credito Diners Club International, Master Card, VISA, American Express;

2) i soci devono contattare direttamente la Segreteria della Società

I cambi di indirizzo di spedizione, e-mail o di qualsiasi altro dato di abbonamento vanno segnalati tempestivamente inviando i nuovi e vecchi dati per posta, fax, e-mail o direttamente sul sito www.minervamedica.it alla sezione "I tuoi abbonamenti - Contatta ufficio abbonamenti"

I reclami per i fascicoli mancanti devono pervenire entro 6 mesi dalla data di pubblicazione del fascicolo

I prezzi dei fascicoli e delle annate arretrati sono disponibili su richiesta

© Edizioni Minerva Medica - Torino 2015

Tutti i diritti sono riservati. Nessuna parte di questa pubblicazione può essere riprodotta, trasmessa e memorizzata in qualsiasi forma e con qualsiasi mezzo. Pubblicazione bimestrale. Autorizzazione del Tribunale di Torino n. 688 del 18-1-1952. Iscrizione nel registro nazionale della stampa di cui alla legge 5-8-1981 n. 416 art. 11 con il numero 00 148 vol. 2 foglio 377 in data 18-8-1982. Pubblicazione periodica mensile - Poste Italiane S.p.A. - Sped. in a.p. - D. L. 353/2003 (conv. in L. 27/02/2004 N° 46) art. 1, comma 1, DCB/CN

Segreteria - Società Italiana di Odontostomatologia e Chirurgia Maxillo-Facciale c/o Università Cattolica del Sacro Cuore, Complesso Integrato Columbus, Via G. Moscati 31-33, 00100 Roma. Tel. 06 3503874 - E-mail: siocmf@hotmail.it
Segretario e Tesoriere Prof. M. Procaccini.



Pubblicazione associata a

La Rivista aderisce al Codice di Autodisciplina degli Editori Medici Scientifici associati a FARMAMEDIA e può essere oggetto di pianificazione pubblicitaria

**XXII Congresso Nazionale
Collegio dei Docenti Universitari
di Discipline Odontostomatologiche**

Milano, 9-11 aprile 2015

Methods. Twenty-two upper incisors were selected from a pool of extracted teeth. The teeth had a single round canal with a long-short cross-section diameter ratio $\leq 2,5$ at 8 mm from the apex and similar root canal anatomy. The teeth were randomly divided into two groups in according to the irrigation technique used: self-adjusting file system (SAF) (group 1) and conventional endodontic needle irrigation (group 2). The groups resulted homogeneous according to the canal width. The same operator performed all experimental procedures. The crowns were sectioned at the cementoenamel junction and the root length was standardized to 13 mm from the apex. The coronal thirds was enlarged with size 1, 2, 3 and 4 Gates-Glidden drill. The working length (WL) was established by the insertion of a 21-mm #10 K-File until its tip appeared at the apical foramen under microscopic vision at 10x. A glide path was performed using size 13, 16 and 19/02 PathFile at 300 rpm. In the group 1 the 1,5-mm-diameter SAF was operated for four minutes and continuous irrigation with 5,25% NaOCl was performed by a VATEA peristaltic pump at a rate of 4 mL/min. In the group 2 the canals was instrumented using Protaper Universal to a size of the F2 instruments at the working length and irrigated with 1 ml 5% NaOCl at every instrument changes. In both groups a final irrigation of 2.0 mL 1% EDTA for 30s was performed. Roots filling was performed with Guttacore Obturators with TopSeal labeled with 0,1wt% Rhodamin B. Transverse sections at 2, 5 and 7 mm from apical foramen were observed using a confocal laser scanning microscopy. Total percentage and maximum depth of sealer penetration were measured and registered using ImageJ software. Statistical analysis was performed by ANOVA and Turkey test.

Results. The 7-mm and 5-mm sections of group 1 showed a significantly higher percentage and maximum depth of sealer penetration respect the group 2. Not significant difference was found at 2-mm sections between the two groups.

Conclusion. The self-adjusting file cleaning-shaping-irrigation system improved the sealer penetration at coronal and middle sections respect the conventional endodontic needle irrigation. At apical sections no differences were found.

HyFlex EDM rotary Ni-Ti prototypes: the effect of an innovative machining technology on Ni-Ti wear

C. Pirani ¹, L. Generali ², F. Iacono ¹, P. Sassatelli ³, C. Prati ¹

¹School of Dentistry, Endodontic Clinical Section, Department of Biomedical and Neuromotor Sciences (DIBINEM), University of Bologna, Alma Mater Studiorum, Bologna, Italy

²Department of Surgery, Medicine, Dentistry and Morphological Sciences with Transplant Surgery, Oncology and Regenerative Medicine Relevance, School of Dentistry, Endodontic Section, University of Modena and Reggio Emilia, Modena, Italy

³Department of Engineering Enzo Ferrari (DIEF), University of Modena and Reggio Emilia, Modena, Italy

Aim. HyFlex EDM files were recently introduced presenting an innovative electro discharge machining (EDM) process of fabrication. The aim of this study was to evaluate the surface and microstructural alterations of new and in vitro used HyFlex EDM Ni-Ti rotary prototypes.

Methods. The surface and microstructural characteristics of 15 new HyFlex EDM instruments were analyzed by ESEM equipped with energy dispersive x-ray spectrophotometry (EDS) and optical metallographic imaging. Instruments were subjected to instrumentation tests on severely curved root canals (ranging between 50° and 70°) of extracted multi-rooted teeth. Once that canal patency was verified with a #10 K-file, the working length was determined by subtracting 1mm. HyFlex EDM files were used with a 16:1 reduction handpiece X-Smart (Dentsply Maillefer, Baillagué, Switzerland) following the manufacture's direction, at 500 rpm and 2,5Ncm, with slightly apical pressure and pecking motion. The operative sequence was: 25/12 at 2/3 of the WL, 10/05 and 25/08 at WL. Irrigation was performed at every change of instrument, with a total amount of 3 ml of 5% NaOCl and 3 ml of 10% EDTA (Ogna, Mugliano, Italy). Each instrument was used in 10 curved canals, washed in an ultrasonic bath containing detergent for 10 min and then autoclaved at 134°C. Surface and microstructural characterizations were repeated on used instruments at same points and with same angulations to compare the pre- and postoperative micrographs, in order to verify the appearance of fractures, unwinding, microcracks, blade disruption and tip deformation.

Results. Surface and microstructural characterization of new instruments revealed the typical features of a NiTi ED-Machined alloy with an irregular and "craters-like" surface. High magnification micrographs disclosed a non-uniform structure were pits, pores and voids caused the peculiar aspect of a "rough-spark-machined" surface. No fractures were registered during instrumentation of curved canals. Surface and microstructural characterization of used files revealed no wear and no degradation of the 25.12 and 25.08 files. The tip segment was confirmed as the most mechanically stressed portion of 10.05 prototypes. All the instruments, after several uses, well-preserved the "craters-like" irregular surface without cutting edge alterations. The metallographic inspection on the cross section of brand new HyFlex EDM files showed an homogeneous martensitic phase. The microstructure appeared uniform from the surface to the bulk, and no microcracks or defect were identified, even at high optical magnification (1000X).

Conclusion. Unaltered spark-machined surface and low microstructural degradation are the main features of recently introduced HyFlex EDM. Caution would be recommended regarding reuse of small HyFlex EDM files. Instruments exhibited a safe in vitro use in presence of severely curved canals.

Vascular endothelia growth factor (VEGF) expression in human tooth germs early and later stage development

F. Mastrangelo ¹, M. T. Sberna ², R. Vinci ², L. Tettamanti ¹, G. Cantatore ², A. Tagliabue ¹, E. F. Gherlone ²

¹Dipartimento di Scienze Chirurgiche e Morfologiche, Università degli Studi dell'Insubria, Varese

²Dipartimento di Scienze Mediche ed Odontostomatologiche, Ateneo Vita e Salute, Università S. Raffaele, Milano

Aim. The angiogenesis plays a crucial role in many human physiological and pathological processes during