This is a pre print version of the following article:

Environmental sustainability of orthopedic devices produced with powder bed fusion / Cappucci, G. M.; Pini, M.; Neri, P.; Marassi, M.; Bassoli, E.; Ferrari, A. M.. - In: JOURNAL OF INDUSTRIAL ECOLOGY. - ISSN 1088-1980. - 24:3(2020), pp. 681-694. [10.1111/jiec.12968]

Terms of use:

The terms and conditions for the reuse of this version of the manuscript are specified in the publishing policy. For all terms of use and more information see the publisher's website.

27/04/2024 16:22

1 Article Type: Research article

2 **Title:** Environmental sustainability of orthopaedic devices produced with powder bed fusion

3 Authors: Cappucci, G. M.^{1,2}, Pini, M.², Neri, P.², Marassi, M.², Bassoli, E.³ & Ferrari, A. M.²

4 Institutions: ¹INSTM - National Interuniversity Consortium of Materials Science and Technology

⁵ ²University of Modena and Reggio Emilia - Department of Sciences and Methods for Engineering

6 ³University of Modena and Reggio Emilia - Department of Engineering "Enzo Ferrari"

Corresponding Author: Cappucci G.M., Piazzale Europa 1, 42124 Reggio Emilia, Italy,
 graziamaria.cappucci@unimore.it, http://www.lcaworkinggroup.unimore.it/site/home.html.

9 **Conflict of Interest Statement:** The authors declare no conflict of interest.

Keywords: life cycle assessment (LCA), product lifetime, industrial ecology, Additive
 manufacturing, Orthopaedic device, Traditional manufacturing

12

13 Abstract:

14 Additive manufacturing consists in melting metallic powders to produce objects from 3D data,

15 layer upon layer. Its industrial applications range from automotive, biomedical (e.g. prosthetic

16 implants for dentistry and orthopedics), aeronautics and others.

17 This study evaluates the possible improvement in environmental performance of laser-based

18 powder bed fusion additive manufacturing systems on prosthetic device production through Life

19 Cycle Assessment (LCA) methodology. Environmental impacts due to manufacturing, use and end

20 of life of the designed solution were assessed. In addition, two powder production technologies, gas

21 atomization (GA) and plasma atomization (PA), were compared in order to establish the most

22 sustainable one. Production via traditional subtractive technologies and the additive manufacturing

23 production were also compared.

3D building was found to have a significant environmental advantage compared to the traditional technology. The powder production process considerably influences on a damage point of view the

additive manufacturing process, however its impact can be mitigated if GA powders are employed.

27 **1 Introduction**

Additive manufacturing (AM) is a 3D building technology that is rapidly increasing among
 manufacturing processes, in which the building process involves layering materials. Its strengths are

its ability to create objects with a high geometrical complexity, which are not possible to obtain in
traditional manufacturing, and the flexibility in meeting customer's requests in terms of design,
without increasing the productive costs.

33 Its aims are perfectly in line with the European Union Industry 4.0 plan (European Parliament 34 Research Service, 2015), which is built on the model of the high-tech strategy of the German 35 government, and whose main objectives are an increased flexibility and productivity in 36 manufacturing, mass customization and better quality.

In order to achieve these ambitious goals a new vision, named the "smart factory", is needed,
including the integration of IT services, such as the digitisation of information and big data analysis,
and of cyber-physical systems, such as embedded sensors, intelligent robots and additive
manufacturing devices.

Additive manufacturing has been designated by the Boston Consulting Group, the worldwide
multinational company in management consulting, as one of the five enabling technologies due to
increased efficiency in material use (Sirkin, Zinser & Rose, 2015).

Powder bed fusion (PBF) is one of the latest terminologies for the designation of an AM process
in which a metal powder layer is laid out over a bed and sintered by a high-energy beam, often a
laser (Gibson, Rosen & Stucker, 2015).

This technology can be applied to a wide range of materials, but is most suited to metals. The opportunity to build metal objects with a complex geometry and high customization potential, which is impossible in traditional manufacturing, is one of the most interesting features from a technological, as well as a business perspective.

51 Selective laser melting (SLM) is one of the commonly used techniques, in which metallic

52 powder is fully melted in high-density and 3D structures (Gibson et al., 2015) rather than sintered,

53 thus giving greater control over material properties such as porosity and crystal structures.

54 Although the technical achievements of AM processes are widely acknowledged, they still need

55 a Life Cycle Assessment (LCA), in order to evaluate the strengths and weaknesses from a

sustainability perspective, in comparison with traditional manufacturing. A literature analysis was
therefore conducted and the main findings are reported below.

58 **1.1 Literature analysis**

As reported by Kellens et al. (2017), regarding AM processes as self-sufficient technologies is
not accurate, as post-processes are often required to reduce surface stresses due to the anisotropy of
AM parts.

62 The same authors provided a wide overview of AM processes compared to the corresponding63 traditional manufacturing processes.

64 For example, Serres, Tidu, Sankare and Hkawka (2011), applied the Eco-Indicator 99

65 methodology (Goedkoop & Spriensma, 2001) to the production of a mechanical component in

66 Ti₆Al₄V alloy, and analyzed the incidence on total damage caused by upstream processes, such as

67 powder production and ingot production, on additive and traditional manufacturing. The authors

showed that the AM involves much lower damage compared to traditional manufacturing, however

69 the two technologies are comparable if larger parts are produced with the AM, due to the

70 considerable amount of metal powder needed to build the component.

71 Peng et al. (2017) applied a system expansion approach to the AM process to model the by-

72 product derived from unmelted loose powder at the end of the productive process. They considered

73 five environmental indicators, global warming potential, acidification potential, Chinese resource

74 depletion potential, eutrophication potential, and respiratory inorganics. They found that an impeller

75 made with titanium alloy totally produced with AM has a higher impact compared to that produced

76 with traditional manufacturing. This environmental damage is mainly due to powder production and

electricity consumption. AM may only have environmental advantages if the impeller is partially

78 produced with traditional manufacturing.

Priarone, Ingarao, Di Lorenzo and Settineri (2016) studied both productive processes (traditional
 and additive manufacturing) from a cradle-to-grave perspective in terms of CO₂ emissions,

81 computed using the carbon emission signature (CES) method proposed by Jeswiet and Kara (2008),

and the energy demand by applying the system expansion with substitution LCI model. They found
that the environmental loads are influenced by the material removal rate - AM is the most
favourable technology when a significant amount of material can be saved, although there is a
higher energy consumption compared to traditional manufacturing when small quantities of
material need to be removed.

Huang et al. (2015) found that AM has a considerable advantage over traditional manufacturing
when different case-studies (EOS, 2013; Krailling & Novi, 2014; Munsch, Wycisk, Kranz, Seyda &
Claus, 2012; the SAVING project, 2009; Tomlin & Meyer, 2011) related to the production of
components for transportation vehicles, are considered and analyzed in order to outline a common
profile.

The use phase plays an important role in damage assessment. Considering a period from 2014 to 2050, AM parts are preferable to the traditional manufactured ones in terms of energy savings, thanks to a significant mass reduction in the components, which entails a lower fuel consumption. Moreover, lower buy-to-fly ratios of AM parts, which were assumed to be 1.5 for all AM processes, in the cradle-to-gate LCI model resulted in lower primary energy use and GHG emissions compared to traditional manufacturing.

In the medical devices production field, the following studies have been published, however
none of them involve a comparison with traditional manufacturing. Baumers, Tuck, Bourell,
Sreenivasan and Hague (2011) analyzed the energy consumptions of two laser sintering platforms
(Sinterstation HiQ+HS and EOSINT P 390) for building two prosthetic parts and found that most
energy is employed for heating and cooling.

103 Sreenivasan, Goel and Bourell (2010) calculated the energy consumption for producing

104 prosthetic parts using polymeric material by defining an energy indicator which enables different

105 selective laser sintering processes to be compared.

106 **1.2** Scope of the LCA study

107 In this study LCA methodology was used to analyse the different levels of impact on the 108 environment of manufacturing hip prostheses using AM and traditional manufacturing processes. In 109 particular, femoral stems produced with Ti_6Al_4V alloy by Powder Bed Fusion technology and by 110 traditional manufacturing, over the whole life cycle, were considered.

111 Due to the relevance of metal powder production in terms of the total damage, gas atomization 112 (GA) and plasma atomization (PA) were compared, in order to evaluate the most sustainable 113 production method.

The advantages of AM compared to traditional manufacturing were also assessed through a social indicator during the impact assessment stage which expresses the acquired utility of the part produced with AM from a social perspective. In this study, the interest in the environmental performance of the product is predominant over its technical performance, although this was taken into account in the environmental analysis in order to provide a result as complete as possible.

119 2 Life cycle of a femoral stem produced with PBF

120 A hip implant is the only effective cure for coxitis, which is a degenerative disease in which the 121 cartilage surrounding the two extremities of the joint, femoral head and the acetabulum,

122 deteriorates.

In its primitive form, coxitis occurs in people over 60, while it can affect younger people due to congenital illness, such as dysplasia (Gruppo Biompianti, 2017). In Europe more than 600,000 hip replacement procedures were performed in 2005 (Kiefer, 2007).

126 The entire life cycle of femoral stems produced with AM was considered taking into account

127 Ti₆Al₄V alloy powder production, femoral stem production, use and end of life phases. The titanium

128 alloy production, titanium alloy powder (40 µm) production with atomization and the production

129 phase with an EOS M290 machine were included. Waste material disposal, such as waste metal

130 recycling and exhausted argon treatment, were also included.

- 131 During the production process, indoor emissions were taken into account, considering PPE
- 132 (personal protective equipment). The main steps in the life cycle of femoral stem production with
- 133 AM are described in Figure 1.

134



Figure 1 System boundaries of femoral stem life cycle with AM

135

136 2.1 Ti₆Al₄V powder production

Ti₆Al₄V powder production is described first considering the GA technology and then the PA.
The main differences between these production processes consist in alloy feeding and atomization
technology.

The PA process uses a Ti_6Al_4V wire feedstock, which is straightened and positioned at the apex of three plasma torches. Each plasma torch provides about 30 kW (Pyrogenesis, 2017) and is fed with argon. Cooling water is fed to each torch and to the atomization tower in order to ensure accurate temperature control. The plasma flow melts the wire, whose droplets solidify into spherical particles when they fall down the atomization tower.

The GA process uses a Ti_6Al_4V bar feedstock that is rotated and, at the same time, lowered into an inductive coil which melts the bar without making contact with it. The melt is then atomized by high-pressure argon jets.

Another important difference between the two technologies is the morphologic atomization efficiency. Morphologic atomization efficiency is the capacity to produce high purity and high sphericity of particles and it is mathematically defined as the ratio of perfectly spherical and pure particles over the total amount of target powder.

PA technology is characterized by nearly 99% morphologic atomization efficiency, while, GA
technology has about 90% morphologic atomization efficiency. These efficiencies are estimated on
the basis of SEM images reported in Popovich, Sufiiarov and Grigoriev (2017).

155 The outgoing argon and powder flows, for both PA and GA technologies, are then separated by 156 the following steps:

• cyclonic separation of Ti₆Al₄V powder from argon;

sieving of Ti₆Al₄V powder in order to accurately separate powder particles with the correct
 particle size distribution and morphology for AM from oversized powder, which is supposed
 to be sold to coating manufactures, and from undersized powder, which is supposed to be
 sent to metal recycling process;

- baghouse filtration which purifies exhausted argon;
- argon recirculation in the atomization process.

The powder produced by PA technologies presents a tap density of 2810 kg/m³ (Advanced
Powder and Coatings Inc., 2017), while powder produced by the GA process has a tap density of
2710 kg/m³ (Venkatesh et al., 2016).

Both atomization processes work 16 hours/day (EOS, 2017) and are characterized by indoor and
local emissions of argon and metals.

169 2.2 Femoral stem production

Femoral stem production takes place in an EOS M290 machine, where fusion is performed by a 400 W laser. The production lasts 61 hours and 21 minutes with a production capacity of 20 femoral stems (Poly-Shape, 2017) per job. After a set-up phase, in which argon is injected in order to minimize the oxygen level, powder is fed by the dispenser system. A 40 µm thick layer is then extended on a titanium plate with a recoater. Laser fusion involves the selective melting of crosssections, previously defined by the CAD model. After each layer has been completely melted, the plate is lowered for a new layer deposition which, in turn, will also be melted.

During the build phase, the argon flow is insufflated in the process chamber in order to prevent the development of an explosive atmosphere due to increase in powder particles and to control the N/O pick-up. An air recirculating filtering system works continuously in order to guarantee the right level of argon purification.

After the job has been completed, the parts are extracted by workers, who wear protective equipment. Extraction involves the separation by sieving solidified parts from the remaining loose powder, which are then reused for the following job. After extraction, the parts are heat treated for two hours at 840° C, cut from the plate with a wire erosion machine and, then, finished with sandblasting and mirror-like polishing. As the parts produced have no internal cavities, depowderization with compressed-air is not considered. Indoor metal emissions are considered, which occur during the part extraction, machine cleaning and cutting of the stems from the building platform. Waste 188 metal powders resulting from machine cleaning and caught by protective equipment are rendered189 inert first and then buried in a residual landfill.

190 **2.3** Use phase

191 The use phase takes into account the surgical stem implantation, a hospital stay for two weeks 192 and medical examinations over the patient's lifetime. The average lifetime of the prosthesis is 193 calculated to be 14.5 years. This value was obtained with a weighted average of current hip joint 194 survivals, which range between 92% at 11 years and 86% at 22 years, as reported by Wyatt, 195 Hooper, Frampton and Rothwell (2014). It is assumed that the first medical examination occurs in 196 the initial weeks after the stem implantation with the second examination occurring in the same 197 year. In normal conditions patients undergo subsequent medical examinations every five years. The 198 medical check-up consists in X-ray examinations, which take 30 minutes, in order to evaluate the 199 effects of wear and tear on the prosthesis. If the patient lives beyond the lifetime of the prosthesis, a 200 surgical removal was considered. Deceases before the stem removal were defined as being equal to 201 25% of total implantations (rate of decease within 10 years from the stem's implantation, 202 Wainwright, Theis, Garneti & Melloh, 2011).

If death occurs before removal, the prosthesis is not removed from the patient, in order topreserve the integrity of the person.

205 2.4 End of life

Femoral stem end of life was defined following direct interviews with technicians from an Italian hospital, the Rizzoli Orthopaedic Institute. These technicians reported that prostheses are surgically removed, sterilized and then archived. No material recycling or prosthesis reuse is performed, according to practices adopted by interviewed technicians.

210 **3** Methods: Life cycle assessment

211 **3.1 Goal and scope definition**

The goal of the study was to assess the environmental impacts of Ti_6Al_4V alloy based femoral stems produced with AM over their entire life cycle in order to identify the environmental hotspots of the system in line with UNI EN ISO 14040-14044 regulations and to propose improvements forimpact mitigation.

216 **3.2** System, functional unit and function of the system

217 The system studied is a bed fusion of Ti_6Al_4V alloy powder. AM is used for the application of

218 biomedical devices, such as femoral stems. Twenty femoral stems produced with AM were

analyzed.

220 **3.3 System boundaries**

221 The system boundaries cover the entire life cycle of the analyzed system ranging from the

 Ti_6Al_4V alloy feedstock and Ti_6Al_4V powder productions to the manufacture, use and end of life

stages of the femoral stems (Figure 1). The system boundaries of the femoral stem production using

traditional manufacturing are shown in Figure 2.

225



241

Figure 2 System boundaries of femoral stems life cycle with traditional manufacturing

242 The production, maintenance and disposal of facilities as well as other auxiliary materials were also included in the present study. Air and indoor emissions as well as solid and liquid waste 243 produced in each step were considered and quantified. The following assumptions were also made: 244 Transport of raw materials, facilities, systems and machines are considered for an average 245 distance of 100 km from the producer to the user, as required by the environmental product 246 247 declaration (EPD) certification, when primary data about distances are not available; Distance of transport of femoral stems from the producer to the final customer is fixed at 248 249 100 km: 40% by rail and 60% by road; 250 Electricity energy production is assumed to be the European mix electricity energy proposed 251 by Ecoinvent database (Ecoinvent Centre, 2018);

Installation of 99.97% efficiency HEPA air filter during femoral stem production and
 powder production steps;

Use of 99.95% efficiency personal protective equipment (filter category P3) during EOS
 M290 machine cleaning, cutting, powder production and exhausted argon treatment steps
 (EOS, 2016).

257 **3.4 Data quality**

Primary data related to the raw materials and to the AM process, the machine characteristics, the consumables needed for the stem production (such as the amount of argon, while the amount of powder is calculated on the basis of other primary data), the post-production treatments were directly collected from a market leader in AM production in Europe.

Another market leader in Europe in prostheses produced by traditional manufacturing was also
interviewed. Use and end of life phases were modeled with secondary data from the literature.
The inventory analysis was modelled in SimaPro 8.5.0 (Pré, 2017) and with vers. 3.4 of the
Ecoinvent database (Ecoinvent Centre, 2018).

266 The LCI model *attributional*, *partitioning*, in terms the presence of the co-product represented by

267 loose powder remaining at the end of the production process, is considered as the most appropriate

to best satisfy the assessment requirements (Pini, Neri & Ferrari, 2018). The use of *substitution* in

the attributional data modelling is not considered adequate, as the co-product is not assumed

identical, due to additional processes to which it has been subjected, to virgin powder, even if theyperform the same function.

The allocation is based on energy criterion, in particular non-renewable and renewable energy consumption is taken into account. Energy allocation is preferred because, from a methodological point of view, it is more representative of the studied system, as it takes into account all the stages in the production of stems, from an energy point of view.

The weight allocation is declined because it would attribute almost all the damage to the coproduct, according to the respective masses involved, neglecting the purpose of the process, that is the production of prostheses. Moreover, this kind of allocation would erroneously equalize the
products from an importance point of view. Allocation based on economic value is not performed
due to the lack of primary data.

281 **3.5 Impact assessment methodology**

The environmental analysis were carried out by the IMPACT 2002+ method, modified in accordance with Pini, Ferrari, Gamberini, Neri and Rimini (2014). Since the IMPACT 2002+ method does not taken into account local and indoor emissions, characterization factors for argon and metal emissions were obtained by adopting a preliminary method (Ferrari et al., 2019) in order to calculate indoor and local human effects. These indicators were introduced in the Life Cycle Impact Assessment (LCIA) method.

The following were thus added to the above mentioned evaluation in order to consider a wider and more representative scenario of the considered system:

- New Carcinogens categories were introduced, Carcinogens indoor and Carcinogens
 local, in particular, new substances are added in the new categories, namely Metals,
 unspecified indoor and Metals, unspecified local with defined characterization factors
 calculated with the method mentioned above.
- 294 In particular, the characterization factor for indoor and local Metals, unspecified result in 295 1642.011 kgC₂H₃Cl eq./kg and 1255.66 kgC₂H₃Cl eq./kg. These values are obtained 296 considering for both factors the damage factor reported in Eco-indicator 99 (EI99) of the analysed substance (6.969E-4 DALY/kg), the fate factor and the population density 297 (namely, $3.13E-5 \text{ m}^2\text{y/m}^3$ and $1.17E-4 \text{ pers/m}^2$, both the fate factor and population 298 299 density belong to Lindane, the substance that in Annex v. 3 of EI99 has a damage factor near to Metals, unspecified), local and indoor fate factors (namely, $7.39E-5 \text{ m}^2\text{y/m}^3$ and 300 1.087E-5 m²v/m³, calculated by Eco-indicator 99 formula considering for local emission 301 an emitting area of 4E8 m² and local concentration calculated by Gaussian Plume 302 303 (Zannetti, 1990), a stationary model used to simulate the air pollutants dispersion into air

304	emitted from a chimney, for indoor emission an emitting area of 25 m ²) and local and
305	indoor population density (considering, namely, 100000 inhabitants for the local area
306	and 2 workers in the shed).

307 A new Non carcinogens category was introduced, Non carcinogens indoor, including • Argon with the calculated damage factor. The limit of argon concentration in a working 308 space, considered to be 500 m³, is equal to 0.18 kg/m^3 and is calculated considering the 309 310 increased percentage of argon (up to 10%) in air. Considering a breath rate of 2.5 m³/h 311 and 8 working hours per day, the indoor argon limit of emission was calculated as 3.57 312 kg. Referring to Europe (with a population density of 386 million, Goedkoop & 313 Spriensma, 2001) and considering an average lifetime of 80 years and a 50 year old man 314 exposed to emissions, the damage factor on human health is 2,18E-6 DALY/kg and the 315 resulting characterization factor is 0.78 kgC₂H₃Cl eq./kg. 316

Social category	Social issues	CF	DAF	NF	WF
	Medical devices	1			
	Suction systems	1			
	Cooling systems	0.5			
	Heating systems	0.6			
Industrial product	Mechanical	0.9		4	0.004
function utility	processings		-1	1	0.001
-	Agricultural machines	0.8			
	Electronic devices	0.6			
	production				
	Movement	0.8			
	transmission				
	Geometry complexity	0.8	-1	1/(0.8+0.8) =	0.001
Product performance	Biocompatibility	0.8		0 625	

317 Table 1 Impact/damage categories added in IMPACT 2002+, with each substance, characterization factor (CF), damage 318 assessment factor (DAF), normalization factor (NF) and weighting factor (WF)

The benefits associated with AM compared to traditional manufacturing were also assessed. The aim was to consider the benefits of an AM product that are not considered by LCIA methods. Two social categories were created: Industrial product function utility and Product performance. The first indicator identifies the field of employment of the stem and the second indicator highlights the technical improvement of the stem produced with AM.

Both consider several new issues that express, from a subjective point of view, positive aspects, and which were introduced in the method with calculated characterization factors. For each social category, characterization factors (CFs), normalization factors (NFs) and weighting factors (WFs) are reported in Table 1.

328 The CF value ranges from 0 to 1, based on shared values with the stakeholders. DAF was set to a

329 value of -1, in order to consider the benefit provided by AM. The NF of the Industrial product

330 function utility is equal to the maximum value of the characterization factors. On the other hand, for

331 Product Performance, the normalization factor is the reverse of the sum of the characterization

factors of its social issues, because the issues can all coexist. WF has a value that is three orders of

magnitude lower than the WF of IMPACT 2002+, in order to prevent an excessive influence on the

and environmental results. Only social issues that are representative of the case study are considered in

- the AM process, which are Medical devices, Geometry complexity and Biocompatibility. A higher
- biocompatibility of the stem produced with AM is possible because of the trabecular structure of the

- 337 surface. This particular geometry, that has been validated from a technical-medical point of view by
- the stakeholders (Castagnini et al., 2018), mimics cellular structures of the bone and is not
- achievable with other manufacturing processes, and leads to an improved osseointegration of the
- 340 prosthesis.

341 **3.6 Life cycle inventory**

342 The most representative data used in the Life Cycle Inventory of 20 femoral stems production343 with the EOS M290 machine with GA powder are reported in Table 2.

344

45	Input	Value	Unit
	Materials		
46	Flooding argon	3.03	kg
217	Building phase argon	25.94	kg
947	Ti ₆ Al ₄ V powder	20.83	kg
48	Energy		
	Electricity	147.26	kWh
49	Transport		
50	Road	6.72	tkm
50	Output	Value	Unit
51	Main product		
	20 femoral stems	1.77	kg
52	Co-product		
52	Loose powder	18.99	kg
33	Indoor emissions		
54	Metals, unspecified indoor	5.95E-9	kg
01	Argon, indoor	1.2E-7	kg
55	Local emissions		_
	Metals, unspecified local	1.9E-3	kg
56	Emissions to air		
57	Metals, unspecified	1.71E-2	kg
57	Argon	2.89E-4	kg
58	Waste to treatment	4 05 0	
	Metal recycling	1.9E-2	kg
59	Disposal to residual landfill of metals captured by filter	2.08E-2	kg

360

Table 2 Inventory input data for the AM process of 20 femoral stems with EOS M290

361 The percentages resulting from the energy allocation between the main product and the co-

362 product were derived from equations (1) and (2):

363 364	$20 \ stems = \frac{1}{(n \ stems)}$	$\frac{n Stems x (NR energy_{1 stem} + R energy_{1 stem})}{tems x (NR energy_{1 stem} + R energy_{1 stem})) + (n kg x (NR energy_{1 kg} + R energy_{1 kg}))} x 100 = 58.32\%$	(1)
365 366			
367	Loose powder =	$=\frac{(n \ kg \ x \ (NR \ energy_{1 \ kg} + R \ energy_{1 \ kg}))}{(n \ Stems \ x \ (NR \ energy_{1 \ stem} + R \ energy_{1 \ stem})) + (n \ kg \ x \ (NR \ energy_{1 \ kg} + R \ energy_{1 \ kg}))} \ x \ 100 = 41.68\%$	(2)
368	where:		

369	• NRenergy _{1stem} is the amount of non-renewable energy, expressed in MJ, required for
370	producing one femoral stem;
371	• Renergy _{1stem} is the amount of renewable energy, expressed in MJ, required for producing
372	one femoral stem;
373	• NRenergy _{1kg} is the amount of non-renewable energy, expressed in MJ, required for
374	producing 1 kg of metallic powder;
375	• Renergy _{1kg} is the amount of renewable energy, expressed in MJ, required for producing
376	1 kg of metallic powder;
377	• n Stems are the number of stems produced in one job;
378	• n kg are the number of kilograms of loose powder remaining at the end of the job.
379	4 Results: Impact assessment
380	An environmental analysis of the life cycle of one femoral stem produced with GA powder was
381	performed. The single score damage was equal to 2.36E-2 Pt ¹ for GA powder usage. The results of
382	the analysis at the mid-point level for GA powder employment are reported in Table 3.
383	Figure S-1 highlights that the most significant contribution to the total damage is due to the
384	Respiratory inorganics impact category (36.34%), which, in turn, is primarily affected by
385	Particulates, $<2.5 \mu m$ (49.94%) due to the production phase (82.72% on total damage of the specific
386	category), in particular for electric energy consumption. Subsequently, the second largest
387	contribution to the total damage is generated by the Non-renewable energy impact category
388	(24.78%), mainly due to Coal, hard (29.42% on total damage of the specific category). This is used
389	in the productive process (78.74% on total damage of the specific substance), especially for energy
390	consumption in primary titanium production, used for the alloy production. In terms of Global
391	warming (24.10%) the main damage is due by Carbon, dioxide fossil (93.02% on total damage of
392	the specific category), especially in the production (68.2% on total damage of the specific

¹ Pt is the abbreviation of "points".

- 393 substance) and use phases (31.77% on total damage of the specific substance), in particular for the
- 394 incineration of hazardous surgery waste.
- 395 The human health is affected by the release of Hydrocarbons, aromatic (80.71%) which
- influence Carcinogens (outdoor environment, 5.7% on total damage of the specific category),
- 397 especially in the use phase (85.96% on total damage of the specific substance) for the production of
- 398 surgery towels in PET.
- 399 The other impact categories provide less than 5% of the total damage.

Impact category	Unit	Total	Production phase	Use phase	End of life
	kg C2H3CI				
Carcinogens	eq	3.41E+00	9.46E-01	2.46E+00	5.64E-04
	kg C2H3Cl				
Non-carcinogens	eq	1.14E+00	8.04E-01	3.33E-01	5.37E-04
Respiratory			=		
inorganics	kg PM2.5 eq	8.70E-02	6.92E-02	1.78E-02	2.92E-05
Ionizing radiation	Bq C-14 eq	9.81E+02	6.37E+02	3.43E+02	2.69E-01
Ozone layer	kg CFC-11				
depletion	eq	5.14E-06	3.75E-06	1.38E-06	2.02E-09
Respiratory organics	kg C2H4 eq	1.69E-02	1.16E-02	5.32E-03	8.43E-06
Aquatic ecotoxicity	kg TEG water	5.55E+03	3.15E+03	2.40E+03	2.61E+00
Terrestrial					
ecotoxicity	kg TEG soil	9.94E+02	6.52E+02	3.41E+02	6.65E-01
Terrestrial acid/nutri	kg SO2 eq	8.19E-01	5.99E-01	2.20E-01	3.27E-04
Land occupation	m2org.arable	5.87E+00	4.31E+00	1.56E+00	1.17E-03
Aquatic acidification	kg SO2 eq	2.51E-01	1.83E-01	6.80E-02	9.68E-05
Aquatic					
eutrophication	kg PO4 P-lim	1.56E-02	1.11E-02	4.55E-03	1.68E-05
Global warming	kg CO2 eq	5.64E+01	3.88E+01	1.76E+01	1.75E-02
Non-renewable					
energy	MJ primary	8.90E+02	5.74E+02	3.16E+02	2.82E-01
Mineral extraction	MJ surplus	8.93E+01	6.39E+01	2.53E+01	1.39E-01
Energia rinnovabile	MJ	9.89E+01	6.93E+01	2.95E+01	5.56E-02
Non-carcinogens,	kg C2H3Cl				
indoor	eq	1.69E-05	1.69E-05	0.00E+00	0.00E+00
Respiratory					
organics, indoor	kg C2H4 eq	0.00E+00	0.00E+00	0.00E+00	0.00E+00
Respiratory					
inorganics, indoor	kg PM2.5 eq	0.00E+00	0.00E+00	0.00E+00	0.00E+00

⁴⁰⁰

Table 3 Characterized LCIA results at mid-point level of 1 femoral stem life cycle with GA powder

401 The endpoint analysis highlights (Table 4) that the phases of the life cycle with the highest

402 environmental burdens are the production (69.32%) and the use phase (30.65%), followed by end of

403 life (0.035%). Moreover, 44.07% of the total damage affects Human Health, 27.26% affects

404 Resources, 24.10% affects Climate Change, 4.75% the Ecosystem Quality, 3.22E-2% the Human

405 health, local and for 3.76E-5% the Human health, indoor. The categories Product performance and

406 Industrial product function utility provide an advantage of -9.85E-2% and -1.23E-1%, respectively.

407

Damage category	Unit	Total	Production phase	Use phase	End of life
Total	Pt	2.36E-02	1.64E-02	7.25E-03	8.37E-06
Human health	Pt	1.04E-02	7.54E-03	2.88E-03	3.32E-06
Resources	Pt	6.45E-03	4.20E-03	2.24E-03	2.77E-06
Climate change	Pt	5.70E-03	3.92E-03	1.78E-03	1.76E-06
Ecosystem quality	Pt	1.12E-03	7.77E-04	3.47E-04	5.11E-07
Human health, local	Pt	7.61E-06	7.61E-06	0.00E+00	0.00E+00
Human health, indoor	Pt	8.89E-09	8.89E-09	0.00E+00	0.00E+00
Product performance	Pt	-2.33E-05	-2.33E-05	0.00E+00	0.00E+00
Industrial product					
function utility	Pt	-2.91E-05	-2.91E-05	0.00E+00	0.00E+00

408

Table 4 LCIA results at end-point level of 1 femoral stem life cycle with GA powder



409 410 411

Figure 3 LCIA results at end-point level of 1 femoral stem AM process with GA powder. Underlying data used to create this figure can be found in supporting information S-2 on the Web

- 412 End-point analysis of one femoral stem production phase is 1.64E-2 Pt, where the AM process
- 413 (86.79%) has the highest environmental load, and post-production treatments (6.46%) and other
- 414 processes (6.73%) contribute to a lesser extent.
- 415 The analysis of the end-point analysis of the AM process (Figure 3) shows that the total damage
- 416 (1.42E-2 Pt) is: 79.88% for Ti₆Al₄V powder production with GA, 6.55% for argon consumption and
- 417 5.18% for electrical energy consumption.
- 418 The damage assessment analysis shows that damage to the Human health accounts for 47.95% of
- 419 the total damage, in particular with the substance Particulates, $<2.5 \mu m$ (air) (49.08%, divided into
- 420 82.54% for powder production and 7.9% for argon consumption).

421 The Resources category provides 24.61% of the total damage, mainly for the substance Coal,

422 hard (35.84%, due especially to the energy production for primary titanium used in alloy powder).

423 The damage to Climate change (24%) is caused almost entirely by the substance Carbon dioxide,

424 fossil (93.51%), 81% emitted for during gas atomization and 6.39% for argon consumption.

425 Aluminium in air affects the category Ecosystem quality (3.75% of the total damage) and is

426 linked to the blasting process for hard coal extraction, used to produce energy, necessary for

427 Ti_6Al_4V bar production process.

Human health, local accounts for 5.24E-2% due almost entirely (99.99%) to Metal, unspecified,
local emitted during parts extraction and machine cleaning.

430 The Human health, indoor category contributes to the total damage with 6.24E-5% due, mainly,

431 to indoor argon emissions during exhausted argon treatment and Ti₆Al₄V powder production, and

432 then to indoor metal emissions occurring while treating exhausted argon, Ti₆Al₄V powder

433 production and femoral stem production processes.

434 Finally, Industrial product function utility and Product performance provide environmental
435 advantages, of -2E-1% and -1.64E-1% respectively.

436 **4.1.1 Comparison of atomization processings**

437 As Ti_6Al_4V powder production causes most of the total damage, a further atomization

438 technology, PA, was investigated in order to assess the most sustainable one. The comparison

439 between 1 kg of Ti₆Al₄V powder produced with GA and PA highlights the higher damage

440 (+12.31%) of PA (2.1E-2 Pt) compared to GA (1.87E-2 Pt). In fact, Ti_6Al_4V powder production

441 with PA provides a higher contribution to the total damage compared to GA because of the greater

442 use of argon (2.56 kg of argon to produce 1 kg of powder) compared to GA (0.007 kg for 1 kg of

- 443 powder), as EOS reported in direct interview (2017), and because of the lower atomization
- 444 productivity of this technology (80 kg of powder produced in 16 hours) compared to GA
- 445 productivity (500 kg in the same cycle) (EOS, 2017).

The damage category with the highest increase is Human health, indoor) which is two orders of
magnitude higher, due to a higher amount of argon sent to treatment, followed by Resources
(+18.38%), Climate change (+11.15%), Human health (+10.6%) and Ecosystem quality (+1.46%).
Moreover, if higher argon consumptions for GA are considered (0.5 kg argon and 2 kg argon for 1
kg of powder), the comparison between the two production technologies provides as result higher
damage for PA (namely, +10.8% and +6.3%).

452 Therefore, as compared with PA, GA, was shown to be the most sustainable option, it was453 chosen for further investigations.

454 **4.1.2** Comparison of femoral stem production lines (traditional versus AM)

455 A comparison between femoral stem production with GA powder and traditional manufacturing456 is reported below.

457 The production phase of 1 femoral stem with traditional manufacturing has a higher impact

458 (2.03E-2 Pt), +24.08% compared to the AM process, caused by the higher rate of metal scraps (15.4

459 kg) that are sent for recycling. Waste powder resulting at the end of the AM process (0.019 kg) and

460 metal scrap resulting from stem's head machining (0.117 kg) are sent to recycling, too.

461 The co-product of AM (about 19 kg of loose powder), in fact, provides a damage reduction to the 462 function. Traditional manufactured parts benefit only from the advantage of the Industrial product function utility. In particular, the indicator Product performance adds to the production of the AM 463 464 stem an advantage (-0,14%) precluded to traditional technology thanks to the novel geometry 465 imprinted by AM, while the benefit derived from Industrial product function utility, considered for 466 traditional production as well, is more limited in the case of additive production (-0.18%) compared 467 to subtractive production (-0.25%) due to allocation of the co-product. As a consequence of the 468 changed geometry, the stem produced with AM, 82.8 g, is lighter (-21%) compared to the one 469 produced with traditional manufacturing, 104.6 g.

A comparison between the complete life cycle damage of the part produced with traditional
manufacturing and AM (with GA powder) is provided (Figure 4). 2.76E-2 Pt is the total damage of

the life cycle of one femoral stem produced with traditional manufacturing which exceeds AM by
16.94% of total damage (2.36E-2 Pt).



477



 $478 \\ 479 \\ 480$

Figure 4 Environmental comparison between life cycle of 1 femoral stem produced with traditional manufacturing and with AM. Underlying data used to create this figure can be found in supporting information S-3 on the Web

481 **4.1.3 Sensitivity analysis**

482 Let us now make a final comparison between the life cycle of one stem made with traditional

483 manufacturing and the life cycle of one reference stem made with additive manufacturing.

484 The reference femoral stem is defined as the stem with the average impact from among 160

485 stems produced in eight jobs.

486 Researchers estimate that loose powder can be reused eight times (Faludi, Baumers, Maskery &

487 Hague, 2016), thus eight jobs, each producing 20 stems, are considered. The first job employes

488 20.83 kg of virgin powder, however the subsequent jobs use the remaining powder from the

489 previous one, adding a small quantity of virgin powder to compensate for the powder lost with

490 waste and printed parts.

491 Damage, in fact, is not constant from one job to another, as the amount of virgin powder 492 introduced into the machine changes and the loose powder coproduct retrieved in each job has a 493 variable impact. In particular, damage decreases until the 7th job, but increases at the 8th job, due to 494 the higher amount of metal powder waste that could not be reused and is sent to recycling. 495 The results of the analysis (Figure S-4) shows that the stem produced traditionally has higher 496 damage compared to the reference stem made with AM (1.81E-2 Pt) of 52.38% which this is due 497 mainly to the reduction of damage associated with the virgin powder introduced in the machine. 498 The LCI modelling of the eight processes (i.e. the eight jobs) is performed once again with 499 attributional, partitioning with energy allocation because loose powder is subjected to further 500 processings, job after job, that could not be adequately expressed with other allocation criterions. 501 5 Conclusions 502 In this work, the environmental sustainability of orthopaedic devices with AM was evaluated 503 with the life cycle assessment methodology. 504 A cradle to grave LCA was applied for one femoral stem produced using AM and GA powder 505 and, as a result, the highest environmental burden was found to be the production phase, followed 506 by the use and end of life phases. 507 The analysis of results highlighted that the main environmental load in the production phase is

due to titanium alloy powder production. The same influence of titanium alloy powder production on total damage was found by Serres et al. (2011) and Peng et al. (2017). In this study two different titanium alloy powder production technologies (GA and PA) were therefore compared in order to highlight the most appropriate option for minimizing environmental loads and protecting human health.

513 The analysis of results illustrates that the most sustainable choice for powder production is GA. 514 An analysis of the benefits derived from the AM process compared to traditional manufacturing 515 was also conducted, taking into account socially positive aspects (never considered before in E-516 LCA studies on additive manufacturing) related to the part produced with AM and concerning the increased biocompatibility and more complex geometry of prostheses. In particular, the indicator
Product performance adds to the life cycle of the AM product an advantage (-0,098%) precluded to
traditional technology, while the benefit derived from Industrial product function utility, considered
for traditional production as well, is more limited in the case of additive production (-0,12%)
compared to subtractive production (-0,18%) due to allocation of the co-product.
These aspects provide an insight into the high level of innovation introduced by this technology,
which is aimed at meeting customer's needs.

Local and indoor emissions were included in the study and their incidence on total damage (namely 3.22E-2% and 3,76E-5%) was found to be very limited, thanks to the high filtration efficiency of HEPA filters and filter mask category P3.

527 The comparison showed that the AM process (in the GA powder usage hypothesis) is the most 528 sustainable option. This is due to the presence of the co-product, represented by loose powder 529 recovered at the end of the productive process, which reduces the damage to the function, choosing 530 energy input as allocation criterion.

A further damage reduction compared to the traditional stem was highlighted when a reference stem, obtained by averaging the impact of 160 stems produced in 8 jobs, is considered. This final analysis highlights the extent of the benefits of additive manufacturing represented by the possibility of reusing loose powder, which is very difficult to investigate without considering all the jobs in which loose powder is employed.

536 6 Acknowledgments

537 This project has received funding from the European Union's Horizon 2020 research 538 and innovation programme under grant agreement No 723699. This publication reflects only the 539 author's view and the Commission is not responsible for any use that may be made of the 540 information it contains.

541 **Funding information**

542 This project has received funding from the European Union's Horizon 2020 research

- and innovation programme under grant agreement No 723699. This publication reflects only the
- 544 author's view and the Commission is not responsible for any use that may be made of the
- 545 information it contains.

546 **References**

- 547 Advanced Powders and Coatings Inc., Titanium alloy powder: Ti-6Al-4V (0-45 mm).
- http://advancedpowders.com/our-plasma-atomized-powders/products/ti-6al-4v-titanium alloy
 powder/#0-45_m. Accessed April, 2017.
- 550 Baumers, M., Tuck, C., Bourell, D. L., Sreenivasan, R., & Hague, R. (2011).
- 551Sustainability of additive manufacturing: measuring the energy consumption of the laser552sintering process. Proceedings of the Institution of Mechanical Engineers, Part B: Journal of553Engineering Manufacture, 225(12), 2228 2239.
- Castagnini, F., Bordini, B., Stea, S., Calderoni, P.P., Masetti, C., & Busanelli, L. (2018). Highly
 porous titanium cup in cementless total hip arthroplasty: registry results at eight years.
 International Orthopaedics (SICOT), 1–7.
- Ecoinvent Center. (2018). *Ecoinvent Database*, version 3.4. Life Cycle Inventories. Retrieved from
 www.ecoinvent.ch. Ecoinvent Association: Zürich. Accessed April, 2018.
- 559 Environmental product declarations. <u>www.environdec.com/</u>. Accessed April, 2017.
- 560 EOS. (2013). EOSCase Study: Light-weight Bracket
- 561 <u>www.eos.info/industries_markets/aerospace/interior</u>. Accessed May, 2017.
- 562 EOS. (2016). EOS M290 machine, Operating Instructions, EOS GmbH.
- 563 EOS. (2017). EOS M290 Datasheet, <u>www.eos.info/m-</u>
 564 <u>solutions/download/datasheet_EOSINT_M290.pdf</u>. Accessed May, 2017.
- EOS. (2017). Personal communication for DREAM project with Nyrhilae, O. General Manager,
 EOS GmbH.
- 567 European commission, Research and Innovation action, number 723699, DREAM.
 568 www.dreameuproject.eu/. Accessed April, 2017.
- European Parliament Research Service. (2015). Industry 4.0: Digitalisation for productivity and
 growth.
- 571www.europarl.europa.eu/RegData/etudes/BRIE/2015/568337/EPRS_BRI(2015)568337_EN572.pdf. Accessed March, 2018.
- 573 European Standard EN149:2001+A1:2009.
- Faludi, J., Baumers, M., Maskery, I. & Hague, R. (2016). Environmental impacts of selective laser
 melting. Do printer, powder, or power dominate? *Journal of Industrial Ecology*, 21(S1),
 S144-S156.
- Ferrari, A. M., Volpi, L., Pini, M., Siligardi, C., García-Muiña, F., & Settembre Blundo, D. (2019).
 Building a Sustainability Benchmarking Framework of Ceramic Tiles Based on Life Cycle
 Sustainability Assessment (LCSA). *Resources*, 8, 11. 10.3390/resources8010011.
- General Programme Instructions for Environmental Product Declarations EPD, the International
 EPD Cooperation, version 1.0. www.environdec.com/en/The-International-
- 582 EPDSystem/General-Programme-Instructions/. Accessed on April, 2017.
- Gibson, I., Rosen, D. & Stucker, B. (2015). Additive Manufacturing Technologies 3D Printing,
 Rapid Prototyping, and Direct Digital Manufacturing . *New York: Springer*.
- Goedkoop, M. & Spriensma, R. (2001). The Eco-indicator 99 A damage oriented method for Life
 cycle Assessment, Methodology Report, Pré Consultants B.V. Amersfoort, 2001.
- 587 Gruppo Bioimpianti, www.bioimpianti.it/biopazienti/?page_id=69. Accessed August 2017.
- Guinée, J.B., Gorrée, M., Heijungs, R., Huppes, G., Kleijn, R., van Oers, L., ... & Huijbregts, M. A.
 J. (2002). *Handbook on Life Cycle Assessment: An Operational Guide to the ISO Standards*.
 Kluwer Academic Publishers, Dordrecht, The Netherlands.
- Huang, R., Riddle, M., Graziano, D., Warren, J., Das, S., Nimbalkar, S., ... & Masanet, E. (2015).
 Energy and emissions saving potential of additive manufacturing: The case of lightweight aircraft components. *Journal of Cleaner Production*, *135*, 1559–1570.
- 594 IPCC. (2007). Climate Change 2007. IPCC Fourth Assessment Report. The Physical Science Basis.
 595 www.ipcc.ch/ipccreports/ar4-wg1.htm. Accessed June, 2018.

- Jeswiet, J. & Kara, S. (2008). Carbon emissions and CESTM in manufacturing. *CIRP Annals— Manufacturing Technology*, 57(1), 17–20.
- Jolliet, O., Margni, M., Charles, R., Humbert, S., Payet, J., Rebitzer, G., & Rosenbaum, R. (2003).
 IMPACT2002+: A new life cycle impact assessment methodology. *The International Journal of Life Cycle Assessment*, 8, 324–330.
- Kellens, K., Baumers, M., Gutowski, T.G., Flanagan, W., Lifset, R., & Duflou, J. R. (2017).
 Environmental Dimensions of Additive Manufacturing, Mapping Application Domains and Their Environmental Implications. *Journal of Industrial Ecology*, 21(S1), S49-S68.
- Kiefer H. (2007). Current Trends in Total Hip Arthroplasty in Europe and Experiences with the
 Bicontact Hip System. In: Sofue M., Endo N. (eds) *Treatment of Osteoarthritic Change in the Hip.* Springer, Tokyo.
- Krailling, G. & Novi, M. (2014). EOS and Airbus Group Innovations Team on Aerospace
 Sustainability Study for Industrial 3D Printing.
 www.eos.info/eos_airbusgroupinnovationteam_aerospace_sustainability_study. Accessed
- June, 2017.
 Munsch, M., Wycisk, E., Kranz, J., Seyda, V., & Claus, E. (2012). Functional products through
 laser additive manufacturing of TiAl6V4. Workshop LAM Laser Additive Manufacturing
 2012, Houston, TX 29.02.2012.
- Peng, S., Li, T., Wang, X., Dong, M., Liu, Z., Shi, J., & Zhang, H. (2017). Toward a sustainable
 impeller production: Environmental comparison among different impeller manufacturing
 methods. *Journal of Industrial Ecology*, 21(S1), S216–S229.
- 617 Pini, M., Ferrari, A. M., Gamberini, R., Neri, P., & Rimini, B. (2014). Life cycle assessment of a
 618 large, thin ceramic tile with advantageous technological properties. *International Journal of*619 *Life Cycle Assessment*, 19(9).
- Pini, M., Neri, P. & Ferrari, A. M. (2018). Environmental performance of waste management in an
 Italian region: how LCI modelling framework could influence the results. *Proceeding of* 25th CIRP Life Cycle Engineering (LCE) Conference. Elsevier B. V.
- Poly-Shape. (2017). Personal communication for DREAM project with Gie, R. Collaborative
 Projects Manager, Poly-Shape.
- Popovich, A.A., Sufiiarov, V. S., & Grigoriev, A.V. (2017). Additive Technologies The Basis of
 Digital Custom Manufacturing. *Industry 4.0. Studies on Entrepreneurship, Structural Change and Industrial Dynamics*, edited by T. Devezas et al., Springer.
- 628 PRé. (2017). *SimaPro 8 Multi user*. PRé Consultants bv, Stationsplein 121, 3818 LE Amersfoort,
 629 The Netherlands.
- Priarone, P., Ingarao, G., Di Lorenzo, R., & Settineri, L. (2017). Influence of material-related
 aspects of additive and subtractive Ti-6Al-4V manufacturing on energy demand and carbon
 dioxide emissions. *Journal of Industrial Ecology*, 21(S1), S191–S202.
- 633 Pyrogenesis Inc., Producing the Most Spherical, Pure Titanium Powders.
- 634 www.pyrogenesis.com/projects/plasma-atomization-production-unit/. Accessed January,
 635 2017.
- 636 Serres, N., Tidu, D., Sankare, S., & Hkawka, F. (2011). Environmental comparison of MESO 637 CLAD process and conventional machining implementing life cycle assessment. *Journal of* 638 *Cleaner Production, 19*, 1117–1124.
- 639 Sirkin, H.L., Zinser, M., & Rose, J. (2015). Why Advanced Manufacturing Will Boost Productivity.
 640 <u>www.bcg.com/publications/2015/lean-and-manufacturing-production-why-advanced-</u>
 641 manufacturing-boost-productivity.aspx. Accessed March, 2018.
- 642 Sreenivasan, R., Goel, A., & Bourell, D. L. (2010). Sustainability issues in laser-based additive
 643 manufacturing, *Physics Procedia*, 5(A), 81-90, ISSN 1875-3892,
- The SAVING project. (2009). Innovation results and R&D case study: Redesigned airline buckle to
 save energy. www.manufacturingthefuture.co.uk/case-studies. Accessed June, 2017.

- Tomlin, M. & Meyer, J. (2011). Topology optimization of an additive layer manufactured (ALM)
 aerospace part. *In Proceedings of the 7th Altair CAE Technology Conference*, 10 May,
 Gaydon, UK.
- 649 UNI EN ISO 14040:2006, www.iso.org/standard/37456.html. Accessed on April, 2017.
- UNI EN ISO 14044:2006, www.iso.org/standard/38498.html. Accessed on April, 2017.
- Venkatesh, V., Pilchak, A. L., Allison, J. L., Ankem, S., Boyer, R., Christodoulou, J.,... &
 Woodfield, A. (2016). *Proceedings of the 13th World Conference on Titanium*, p.147. John
 Wiley & Sons.
- Wainwright, C., Theis, J. C., Garneti, N., & Melloh, M. (2011). Age at hip or knee joint
 replacement surgery predicts likelihood of revision surgery. *The British Editorial Society of Bone & Joint Surgery*, 93(10), 1411-5.
- Wyatt, M., Hooper, G., Frampton, C., & Rothwell, A. (2014). Survival outcomes of cemented
 compared to uncemented stems in primary total hip replacement. *World Journal of Orthopaedics*, 5(5), 591–596.
- Zannetti, P. (1990). Gaussian models. In: *Air Pollution Modeling*. Springer, Boston, MA, USA, pp.
 141–183.
- 662

663 Supporting information

- 664 Supporting information 1: This supporting information provides data about the weighted results by
- 665 impact categories of 1 femoral stem life cycle produced with AM and with GA powder. Underlying
- data used to create this figure can be found in supporting information S-5 on the Web
- 667 Supporting information 2: This supporting information provides data plotted in figure 3 of the main
- 668 text.
- 669 Supporting information 3: This supporting information provides the data used in figure 4 of the
- 670 main text.
- 671 Supporting information 4: This supporting information provides data about the comparison between
- endpoint results of 1-cycle-approach (comparison between one traditional stem and one AM stem,
- top histogram) and 8-jobs-approach (comparison between one traditional stem and one reference
- stem (AM), bottom histogram). Underlying data used to create this figure can be found in
- 675 supporting information S-6 (top histogram) and S-3 (bottom histogram)
- 676 Supporting information 5: This supporting information provides data plotted in figure S-1 of the
- 677 Supporting information.
- 678 Supporting information 6: This supporting information provides data about the underlying data used
- 679 for the comparison between results of 1-cycle-approach (top histogram of figure S-4)

680 Figure Legends

- 681 Figure 1: System boundaries of femoral stem life cycle with AM
- 682 Figure 2: System boundaries of femoral stems life cycle with traditional manufacturing
- 683 Figure 3: LCIA results at end-point level of 1 femoral stem AM process with GA powder.
- 684 Underlying data used to create this figure can be found in supporting information S-2 on the Web
- 685 Figure 4: Environmental comparison between life cycle of 1 femoral stem produced with traditional
- 686 manufacturing and with AM. Underlying data used to create this figure can be found in supporting
- 687 information S-3 on the Web